EXTERNAL



ADQCC – EMIRATES METROLOGY INSTITUTE (EMI)

Establishing a Force and Torque Laboratory that meets the requirements of ISO 17025

Ray Jenkins

T&M 2015 Conference, 13 October 2015 Lord Charles Hotel, Somerset West, Cape Town, RSA



EMI laboratories are located in the CERT Health Science Building



Pressure and Vacuum

* Covered by Initial Assessment

Check List for ISO 17025

• EMI has produced a 68 page document to be used as a check list for its own Internal Audits. This is based on the NIST document that is available:

http://www.nist.gov/pml/wmd/labmetrology/upload/ISOIEC17025Cklist-2009-compatibleWord2003-2.doc

• EMI chose the United Kingdom Accreditation Service – UKAS – as the accreditation body for assessment to ISO 17025 (Calibration). UKAS is a member of the International Laboratory Accreditation Cooperation (ILAC) and a signatory of the ILAC Mutual Recognition Arrangement (MRA)

Main UKAS Activities – Lead Assessor

• Quality Manual - Reviewed prior to visit

• Organisation

Legal Status, Resources, Organisation Structure, Responsibility and Authority, Independence, Impartiality, Integrity and Confidentiality

Management System

Control of documents Control of records Review of requests, tenders and contracts Sub-contracting of calibrations Purchasing services and supplies Service to the customer Complaints Control of non-conforming work Corrective and preventive action process Internal Audit Management review Supervision and monitoring of staff

Main UKAS Activities – Technical Assessor

Technical Competence requirements included:

- Personnel
- Contract review implementation
- Reporting the results (proposed example certificates)
- Accommodation and environmental conditions
- Calibration Methods
- Uncertainty of measurement
- Control of data
- Equipment
- Measurement traceability arrangements
- Handling of calibration items
- Assuring the quality of calibration results

5 MN Force Standard Machine



- The Force Standard Machine (FSM) uses reference force transducers calibrated in an FSM of an NMI that has an appropriate CMC (PTB)
- Forces can be applied in compression and tension, both incrementally and decrementally, over the range 50 kN to 5000 kN
- The machine can be operated under manual control or computer control and store readings from a range of different indicators (DMP39, DMP40, etc.) on a data file
- The uncertainty of the applied force is provisionally 0.02 %

1000 N·m Torque Standard Machine



(AD)

1000 N·m Torque Standard Machine



- Torques are applied by the manual application of weights to scalepans attached to a lever by strip hinges
- The lever itself is supported on a strip hinge arrangement
- All of the strip hinges have strain gauges bonded to both sides so that they are effectively bending moment transducers having a high sensitivity
- The lever is balanced by bringing the summed output of the strain gauge bridges to zero

1000 N·m Torque Standard Machine

- Torques can be applied in the clockwise and anticlockwise directions, both incrementally and decrementally, over the range 0.5 N·m to 1000 N·m
- The weights are applied and removed by hand and are located on the scalepans, and on each other, using a conical location
- The free movement of the scalepans and the beam itself is damped using an Eddy current damping arrangement
- The machine is operated using a computer with manual entry of data or with data entry through an inbuilt mV/V indicator
- Columns under both sides of the lever restrict its movement to avoid overloading the sensitive strain-gauged strip hinges
- The lever is brought to a position of balance using a manually-driven high-ratio gearbox
- The uncertainty is estimated as 0.01 % of applied torque, and is being re-evaluated at the lower values of torque

Control computer displays for Force and Torque Standard Machines



Force resolution is 0.001 kN; Torque resolution is 0.00001 N·m

5 MN Force Standard Machine -Procedures

- **ISO376h** for the calibration of Force Proving Devices with ten equal force steps, applied both incrementally and decrementally, and with a creep and creep recovery test at the end of the calibration. Positions of rotation are 0°, 120° and 240°
- **Repeatability Test** to determine the repeatability of the system at the forces used in the ISO376h procedure. A total of seven runs are made in the 0° position only

• **High Accuracy Test** for the evaluation of Force Transfer Standards used for the verification of Force Standard Machines. The test gives equal weighting to increasing and decreasing forces and uses positions of rotation of 0°, 90°, 180° and 270°. Forces are applied in steps of 10 %, 20 % by 20 % to 100 % of maximum force. The procedure includes a 30 min creep and creep recovery test and a test to measure the temperature coefficient of span over at least 2 °C

It is designed to provide traceability for the EMI 5 MN FSM

1000 N·m Torque Standard Machine -Procedures

• EURAMET Guide cg-14 is used for the calibration of Static Torque Measuring Devices. Calibrations can also be made in accordance with DIN 51309 and BS 7882

• Inter Laboratory Comparison Test is based on cg-14 for the ILC with Norbar Torque Tools, Banbury, UK. The intercomparison covers clockwise and anticlockwise torque, for increasing values only and includes supported and unsupported beams at Norbar

• **Hinge Moment Calibration** is undertaken using small weights applied to the lever in its free condition. Calibration is also undertaken at 1000 N·m using a 1000 N·m reference torque transducer

• Sensitivity Test is undertaken with the lever in its free condition and with a torque of 1000 N·m applied to determine the hysteresis in the strain-gauged strip hinges

• Equality of Lever Length is determined by the application of a series of equal forces to both scalepans, starting with zero force applied

Calibration Results Sheet

- The Calibration Results Sheet (CRS) is used to record information by hand during a calibration (Date, observer, ambient conditions and device temperature). The sheet can be used to record the results obtained during a calibration if necessary, but the calibration results are normally either recorded automatically by the computer, or are entered by hand using the computer keyboard
- It is important to use a CRS when using the 1000 N·m Torque Standard Machine, as it guides the operator through the procedure of applying the weights by hand and balancing the lever
- Data from the CRS is entered on the associated Excel Calibration Analysis Sheet that analyses all the information and produces the results in a form that can be transferred to the Calibration Certificate
- It is planned to use a second computer with the Force and Torque Standard Machines to avoid the need to use a CRS and to also record data from devices which cannot be interfaced with the software provided

Calibration Results Sheet for ISO 376

	CALIBR	ATION RES	ULTS SHE	ET (ENVIRO	ONMENTA	L) FOR FO	RCE TRAN	SDUCER		
				FORIS	O 376					
Date:			Me	etrologist:						
Work Order Number:										
Data file reference:										
	Colibrati	on mode:				cort				
Transdu	canorada	I number:								
Transd	ucer seria	number:				1				
Indica	ator mode	number:)				
Indic	ator seria	number:)	Check instruments are in			
	ndicator re	solution:)				
Environr	ment sens	or model:)				
Environ	ment sen	sor serial:)	Cambration			
Device t	emperatu	re model:)				
Device	temperati	ure serial:)				
Humidity / %rh										
Temperature / °C										
Pressure / hPa										
Device / °C						-				
		Tests X ₁ X ₂			Tests X ₃ X ₄			Tests X ₅ X ₆		
					Humidity / %rh					
					Temperature / °C					
					Pressure / hPa					
					Device / °C					
Sheet checked by:					Name					
					Signature					
					Signature	•				
Transfer of Data to CAS checked by:					Name					
					Signature					

(AD)

Calibration Analysis Sheet - Classification

Limits for Classification / %														
Class	0.05	0.1	0.2	0.5	1	2	5							
b	0.05	0.10	0.20	0.50	1.00	2.00	5.00							
b'	0.025	0.05	0.10	0.25	0.50	1.00	2.50							
f _a	0.025	0.05	0.10	0.25	0.50	1.00	2.50							
f_0	0.0125	0.025	0.05	0.125	0.25	0.50	1.25							
h	0.063	0.125	0.25	0.63	1.25	2.50	6.25							
r	0.025	0.05	0.10	0.25	0.50	1.00	2.50		Torque	Class				
									N∙m					
1000	0.2							Min	1000	0.2				
b	10	10	0.2	0.5	1	2	5	0.2	800	0.2				
b'	0.05	0.1	0.2	0.5	1	2	5	0.05	600	0.2				
f _a	0.05	0.1	0.2	0.5	1	2	5	0.05	400	0.2				
f_0	0.05	0.1	0.2	0.5	1	2	5	0.05	200	0.2				
h	0.05	0.1	0.2	0.5	1	2	5	0.05	100	0.2				
r	0.05	0.1	0.2	0.5	1	2	5	0.05						
800	0.2							Min						
b	10	10	0.2	0.5	1	2	5	0.2						
b'	0.05	0.1	0.2	0.5	1	2	5	0.05	NB	A classification must extend down to at				
f a	0.05	0.1	0.2	0.5	1	2	5	0.05		least 50 % of the range.				
f_0	0.05	0.1	0.2	0.5	1	2	5	0.05		Class 10 means that no classification is				
h	0.05	0.1	0.2	0.5	1	2	5	0.05		possible to EURAMET cg-14.				
r	0.05	0.1	0.2	0.5	1	2	5	0.05						

Calibration Analysis Sheet - Uncertainty



Storage of Data

Data is filed and stored on the EMI "Sharepoint" main server using the conventional method of using a series of folders:

□ Manufacturer

□ Model number / capacity

□ Serial number

□ Date of test(s) (ISO format)

Under the Date of test(s) folder, the data file is stored with a file name:

Serial number-Date of test-Mode-Procedure/Macro-Comment eg:

58622-20150625-c-ISO376h

For force transducer 58622 calibrated on 25 June 2015 in compression to ISO 376 for increasing and decreasing forces. If a second calibration is made to the same procedure, on the same day, the letter A is added immediately after the ISO date

Storage of Data

For torque transducer 59737, calibrated at Norbar on 20 July 2015 in the anticlockwise direction to the procedure F005A (for the ILC with Norbar Torque Tools), the data file would be:

59737-20150720-acw-F005A-Norbar unsupported

As well as containing the data file, the Date of test(s) folder contains:

- Scanned copy of the appropriate Calibration Results Sheet
- The associated Excel Calibration Analysis Sheet that analyses the calibration results
- Calibration certificate in Word format
- Scanned copy of the signed and stamped calibration certificate

We are aiming for a "paperless" system in the future

5 MN FSM Repeatability Tests 58622

5 MN FSM Repeatability Tests 61894

5 MN FSM Repeatability Tests 58671

Calibrations to ISO 376

ΡΤΒ

LNE

Intercomparison with Norbar

Characteristic Graphs for Torque Transducer

Results of Intercomparison with Norbar

Conclusions

- It is essential to establish the protocol for maintaining traceability in the long-term, using appropriate calibration procedures which can be undertaken in-house, or by an acceptable external calibration laboratory
- A regular monitoring program is needed for all the important elements of a Force or Torque Standard Machine
- Participation in Inter Laboratory Comparisons is also essential
- The routine calibration process should be automated as much as possible – Data collection, Data analysis and Certificate preparation
- Spreadsheets should give graphical information to assist in easily identifying anomalies, and must be validated carefully

